

Notice

(Inv.No.OR 61-2023-31R)

Dear Sirs;

Referring to the A/M invitation, We would like to attract your attention to the following amendments :

- amend closing date will be 14-9-2023, instead of 6-9-2023.
- adding the technical descriptions of (Endoscopic Video System for mediastinoscope) to the technical descriptions of (Video mediastinoscope),according to the attached schedule.
- amend the appraisal cost to be : (62,500)\$ (sixty two thousand five hundred USD)which represent the A/M equipment.
- Amount of warranty & maintenance is (13,991\$)(Thirteen thousand & nine hundred ninty one USD) for five years including first year that presented F.O.C by manufacturing Co.
- amend bid bond to be : 6,250.00\$ (six thousand ,two hundred fifty USD
- inside training to be for one week on account of the second party.

THANKS FOR YOUR COOPERATION

-

م/ تنويه

الدعوة المرقمة: 61-2023-31R

تحية طيبة

- اشارة الى الدعوة موضوعه البحث انفا ، نود اعلامكم بانها قد تمت اضافة التعديلات التالية :
- تعديل موعد الغلق ليكون ١٤-٩-٢٠٢٣ بدلا من ٦-٩-٢٠٢٣.
 - اضافة المواصفات الفنية للجهاز (Endoscopic Video System for mediastinoscope) الى المواصفات الفنية للجهاز (Video mediastinoscope) و حسب الجدول الفني المرفق .
 - تعديل الكلفة التخمينية لتكون (62,500) دولار (اثنان و ستون الف و خمسمائة دولار)، لتمثل الأجهزة انفا .
 - مبلغ الضمان والصيانة لمدة خمسة سنوات من ضمنها السنة الأولى التي تقدم مجاناً من الشركة المصنعة للجهاز الواحد هو (13,991USD) ثلاثة عشر الف وتسعمائة وواحد وتسعون دولار فقط.
 - تعديل التامينات الأولية لتكون : 6,250.00 دولار (ستة الاف و مئتان وخمسون دولار)
 - التدريب الموقعي لمدة اسبوع و على حساب الطرف الثاني.

شاكرين لكم تعاونكم معنا ...

Application: Cardio- pulmonary Surgery

<u>VIDEO PROCESSOR</u>	
Resolution	HD or better
Advanced imaging techniques	Yes, Multi- model image
Camera inputs	≥1
Controls/features	Yes
Hue adjustment	Yes
Illumination	Yes, Automatic, manual
White balance	Automatic or , manual
Video gain	Yes
Sharpness	Yes
	Zoom , freeze
Outputs	NTSC ,PAL, Digital
	USB or VGA or Ethernet
Pixel resolution	1920 x 1080 or more
VIDEO CAMERA	HD Or better
LIGHT SOURCE	

Lamp type	LED
Lamp life, hr	≥30,000 for LED ,
Display	≥26" LED or better
Medical Cart	YES, for full system
Suction unit	Yes with tubing set and Canister 1L or more
Power & Environmental Requirements	<p>Power input to be 220-240VAC, 50Hz.</p> <p>The supplier should be confirmed that: The Equipment is suitable for work in climate conditions of Iraq in terms of temperature and humidity</p>