

المواصفات الخاصة بعدة فحص الكلور

- ١-تتص المواصفة القياسية العراقية الخاصة بمياه الشرب المرقمة ٤١٧ بانه لا تقل تركيز الكلور المتبقي لعينات المياه عن (٠.٣) ملغم / لتر ولا تزيد عن (٥) ملغم / لتر
- ٢-قيمة الدالة الحامضية لمياه الاسالة وحسب الموافقة القياسية من ٦.٥-٨.٥
- ٣-تكون عدة فحص الكلور على شكل جهاز حديث رقمي (Portable) لضمان نتائج اكثر دقة علما ان هذا الجهاز مرفق معه محاليل قياسية لضمان دقة في الجهاز
- ٤-يتم ارسال واحد فقط لاغراض التقييم مع ملحقاته من المحاليل القياسية مع Reagents

Specifications for chlorine testing kit

- 1-The Iraqi standard for drinking water numbered 417 includes that the concentration of residual chlorine for water samples is not less than 0.3 mg / liter and not more than 5 mg / liter
- 2-The value of the acidity function of liquefied water according to the standard specification from 6.5-8.5
- 3-The chlorine test kit is in the form of a modern digital device (portable) to ensure more accurate results, knowing that this device is attached with standard solutions to ensure accuracy in the device
- 4-Only one device is sent for evaluation purposes with its accessories of standard solutions with reagents