

CODE NO:

Device Name: automated blood components
Extractor (separator)

DATE:

APPLICATION	Used to process centrifuged whole blood to components(red blood cells ,plasmaand platelet)
FDA CLEARANCER OR CE MARK	YES
Type	Automated
Technology	Top & (Top & Bottom)
Built in scales	Yes
Where used	Main blood bank
Configuration	bench type
Through put capacity / bags	≥ 1 bag
Optical sensors	Yes , built in
Bags type	Different type of blood bags
Data management	Yes , LAN net work
Barcode scanner	Yes

CODE NO:

Device Name: automated blood components
Extractor (separator)

DATE:

Sealing and stripping	Yes automatic
UPS	YES , ≥ 15 min
Display	Yes , color touch screen OR touch pad
Power supply	220/240VAC, 50/60, three pin (G type) plug
Environmental requirements	the supplier confirms the equipment is suitable for working in the climate conditions of Iraq in terms of temperature & humidity .