

CODE NO:

DEVICE NAME: Deep Freezer

DATE:2023

| | |
|--|--|
| Application | Used for freezing of plasma and related blood component. |
| CONFIGURATION | Upright type |
| FREEZANG CAPACITY , L | 400 to 500 L |
| SHELVES / DRAWERS | YES |
| <ul style="list-style-type: none"> • Number • Type | ≥ 3 Stainless steel shelves or better, adjustable. |
| DOORS | <ul style="list-style-type: none"> • Single door or double • lockable |
| CASTERS | Moving casters with locks |
| CHAMBERS MATERIL | <ul style="list-style-type: none"> • inner chamber Stainless steel • outer chamber heavy duty material |
| INTERIOR LIGHT | Yes, high quality LED side lights |
| Defrost | YES |
| TEMPRETURE CONTROL , °C | Up to -80 °C , adjustable |
| AMBIENT CONDITION temp. | 5°C to 40°C |
| COMPRESSOR | Yes |
| Micro control | Touch screen |

| | | |
|--------------------------------------|---|----------------|
| توقيع مدير قسم ادارة الاجهزة الطبية/ | توقيع مدير شعبة الاجهزة الطبية الخدمية/ | توقيع المهندس/ |
| مصادقة اللجنة الاستشارية | | |
| توقيع الدكتور/ | توقيع الدكتور/ | توقيع الدكتور/ |

CODE NO:

DEVICE NAME: Deep Freezer

DATE:2023

| | |
|---|--|
| Control – ALARMS SYSTEM <ul style="list-style-type: none"> • Alarms • Temp indicator • Chart Recorder | Acoustic & visual alarms for High / low temperature, power failure, Ajar door , low battery or others. Digital LCD display. Yes. |
| DATA PORTS | USB/SD card |
| Battery backup | Yes . |
| POWER REQUIRMENTS | 220-240V, 50-60Hz. |
| ENVIROMENTAL REQUIRMENTS | The equipment suitable for working in the climate conditions of Iraq in terms of temperature & humidity |

| | | |
|--------------------------------------|---|----------------|
| توقيع مدير قسم ادارة الاجهزة الطبية/ | توقيع مدير شعبة الاجهزة الطبية الخدمية/ | توقيع المهندس/ |
| مصادقة اللجنة الاستشارية | | |
| توقيع الدكتور/ | توقيع الدكتور/ | توقيع الدكتور/ |