

CODE NO:

DEVICE NAME : gel card incubator

DATE:2024

- **Purpose :** used in blood bank laboratory that matched with our gel card centrifuge
- **Configuration:** tabletop
- **Inner chamber stainless-steel and out chamber heavy duty material.**
- **FDA or CE Clearance :** yes
- 2 independent incubation zones with a capacity for 10 Gel cards
- **Capacity:**  $\geq 10$  cards
- **Temperature Range :**  $37^{\circ}\text{C}$
- **Temperature Accuracy :**  $\leq \pm 1^{\circ}\text{C}$
- **Temperature control:** yes ,programmable time temperature
- **Time Range:** 0-99min59s
- **Alarm :** yes ,emergency stop
- **LCD INDICATOR :** Temperature & time
- **Display:** yes
- **Power requirements:** 220 – 240 VAC, 50- 60 HZ
- **Environmental requirements:** the equipment suitable for working in the climate conditions of Iraq in terms of temperature & humidity

توقيع مدير قسم ادارة الاجهزة الطبية/	توقيع مدير الشعبة الفنية/	توقيع المهندس/
مصادقة اللجنة الاستشارية		
توقيع الدكتور/	توقيع الدكتور/	توقيع الدكتور/