

CODE NO:

**DEVICE NAME: Portable fridge cooler box**  
**(blood bags)**

DATE:2023

<b>Application</b>	Used for transportation of blood & blood component bags.
<b>CONFIGURATION</b>	Chest type
<b>CAPACITY</b>	50 bags of 450ml
<b>TEMPRETURE CONTROL , °C</b>	4°C,+ 2 °C , adjustable
<b>DOORS</b>	<ul style="list-style-type: none"> <li>• Single door</li> <li>• lockable</li> </ul>
<b>CASTERS</b>	Moving casters with locks
<b>CHAMBERS MATERIL</b>	<ul style="list-style-type: none"> <li>• inner chamber Stainless steel</li> <li>• outer chamber polyurethane</li> </ul>
<b>INTERIOR LIGHT</b>	Yes
<b>Display</b>	Digital display
<b>COMPRESSOR</b>	Yes
<b>Control</b>	Microprocessor
<b>Control – ALARMS SYSTEM</b>	
<ul style="list-style-type: none"> <li>• Alarms</li> </ul>	Acoustic & visual alarms for High / low temperature, power failure.
<ul style="list-style-type: none"> <li>• Temp indicator</li> </ul>	Digital LCD display.
<ul style="list-style-type: none"> <li>• Chart Recorder</li> </ul>	Yes.
<b>DATA PORTS</b>	USB/SD card

توقيع مدير قسم ادارة الاجهزة الطبية/	توقيع مدير شعبة الاجهزة الطبية الخدمية/	توقيع المهندس/
مصادقة اللجنة الاستشارية		
توقيع الدكتور/	توقيع الدكتور/	توقيع الدكتور/

CODE NO:

**DEVICE NAME: Portable fridge cooler box**  
(blood bags)

DATE:2023

Battery backup	Yes .
POWER REQUIRMENTS	220-240V, 50-60Hz.
ENVIROMENTAL REQUIRMENTS	The equipment suitable for working in the climate conditions of Iraq in terms of temperature & humidity

توقيع مدير قسم ادارة الاجهزة الطبية/	توقيع مدير شعبة الاجهزة الطبية الخدمية/	توقيع المهندس/
مصادقة اللجنة الاستشارية		
توقيع الدكتور/	توقيع الدكتور/	توقيع الدكتور/